

Material Safety Data Sheet

Gold Latex Powder Free Examination Gloves

Physical Data - Composition

Natural Rubber Latex	≥ 90%
Sulphur	< 1.5%
Zinc Oxide Active	< 0.5%
Accelerator	< 1.0%
Stabilizer	< 0.2%
Titanium Dioxide	< 0.8%
Anti-Oxidant	< 0.3%
Dispersing Agent	< 0.1%
Anti-Ozonant	< 0.8%

Fire & Explosion

Stability: Stable

Extinguishing Media: Foam, Carbon Dioxide, and Water

Precaution

This product contains natural rubber latex which may cause allergic reactions. Other components used in making gloves may also cause allergic reactions in some individuals. If user or a patient is allergic to latex or experience any discomfort, discontinue use immediately and consult with a physician. Do not reuse gloves. Good quality exam gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.

Storage

Store in original packing in a cool, dry and well ventilated area, away from dust, sunlight, moisture, X-ray, and excessive heat above 100°F (37°C).

DISCLAIMER OF EXPRESSED AND IMPLIED WARRANTIES

Although reasonable care has been taken in the preparation of this document, we extend no warranties and make no representations as to the accuracy or completeness of the information contained therein, and assume no responsibility regarding the suitability of this information for the user's intended purposes or for the consequences of its use. Individuals should make a determination as to the suitability of the information for their particular purposes(s).

VR:08/11 KS